

Cellcorder® Spike Probe User's Guide



7775 West Oakland Park Blvd
Sunrise, FL 33351

Tel: (954) 377-7101 Fax: (954) 377-7042

www.alber.com

4200-060 Rev 2.01

Cellcorder® Spike Probe User's Guide



7775 West Oakland Park Blvd

Sunrise, FL 33351

Tel: (954) 377-7101 Fax: (954) 377-7042

www.alber.com

Table of Contents

1	DEFINITIONS.....	1
2	GENERAL DESCRIPTION	1
3	RESISTANCE MEASUREMENT USING SPIKE PROBES	3
3.1	Make Contact To The Battery Post	4
3.2	Apply Firm Pressure	4
3.3	Do Not Move The Probes.....	4
3.4	Inaccessible Battery Posts	5
4	CRT RESISTANCE TESTING	5
4.1	CLC-200 Resistance Testing.....	6
5	VOLTAGE TESTING	6
6	HOW TO MAKE SPIKE PROBE CONNECTIONS	6
6.1	Combined and Computed Readings	6
6.1.1	Combined Internal Resistance.....	7
6.1.1.1	Measuring Cell Resistance	7
6.1.1.2	Measuring Cell Plus Intercell Resistance	8
6.1.2	Computed Internal Resistance.....	8
6.1.2.1	Measuring Computed Cell Plus Intercell Resistance.....	8
6.1.2.2	Measuring Resistance with Intercell Probes.....	9
7	INDEX	10

Table of Figures

Figure 1. Cellcorder Models, CRT & CLC-200.....	1
Figure 2. Coaxial Tip	2
Figure 3. The Straight-line Probe Design, Detachable Probes	2
Figure 4. Table Of Spike Probe Options	3
Figure 5. Table Of Handle Set Options	3
Figure 6. Table of Probe Test Functions And Conditions	5
Figure 7. Measuring Cell Resistance.....	7
Figure 8. Combined Cell and Intercell Resistance	8
Figure 9. Computed Internal Resistance Step 1 & 2	8
Figure 10. Cell And Intercell Resistance Measurement	9

1 Definitions

Following are a few definitions of terms used in this manual.

Battery	Two or more cells connected together electrically in series, parallel or both. See also <i>Cell</i> .
Cell	The basic electrochemical unit that receives, stores, and delivers electrical energy. Often referred to as a <i>Battery</i> .
Jar or Module	The container that holds a cell or group of cells.
Terminal or Post	The part of a cell to which the external circuit is connected.

2 General Description

This manual describes the Spike Probes (Test Cables) for Albér Cellcorders. Refer to your Cellcorder User's Guide for complete descriptions of Cellcorder menus and voltage and resistance testing. The spike probes differ with respect to the types of connectors used to attach the probes to the various models of Cellcorders.



Figure 1. Cellcorder Models, CRT & CLC-200

The Cellcorder Spike Probes are for use with hard-to-access battery terminal posts. Using these probes, it is possible to measure the internal resistance of a cell or module and take float voltage readings. The coaxial design of the probe tips ensures the current and voltage pins make solid contact with the battery posts. Depending on the Cellcorder model, the probes can also be used for combined or computed resistance readings.

The probes are designed for 6 volt and 12 volt UPS cabinet applications where access is both difficult and safety is a concern. However, the probes may also be used on any cell or module up to 12V.


When used with the Cellcorder, the LEDs on the probes light the work area and indicate test status. While the LED-lighted probes are compatible with most Cellcorders, some early models require a factory-installed retrofit. Contact Alber for details. The Test push button on the red (positive) probe lets the technician start the test away from the Cellcorder.

NOTE:

The LEDs do not light when used with the CLC-200.

Modular Design

Each probe assembly has a two piece design that consists of a 7" handle set with an 8 ft. lead, and a detachable probe with a coaxial tip. The probe is a straight line or 45° angled design in various lengths, thus making more locations safely accessible. It terminates in a 5/8" diameter coaxial copper tip.



To calculate resistance, the outer portion of the probe tip conducts test current and the center of the tip senses cell/module voltage. The LED is mounted on the nonconductive PVC housing just above the tip.

Figure 2. Coaxial Tip



Figure 3. The Straight-line Probe Design, Detachable Probes

Figure 4. Table Of Spike Probe Options shows the detachable probe options. Overall length is the probe plus handle and does not include the 8' lead that connects the probe to the Cellcorder.

Detachable Probes

NOTE:

Part numbers with '-45' indicate a 45° angled probe.

Spike Probe Options		
Check www.alber.com for the latest models.		
Probe Part Number	Probe "x" Dimension	Overall Length

1101-008	4.75"	11.75"
1101-008-45	6.00"	13.00"
1101-006	9.25"	16.25"
1101-006-45	10.38"	17.38"
1101-007	23.25"	30.25"
1101-007-45	24.38"	31.38"

Figure 4. Table Of Spike Probe Options

Handle Sets

The 7" handle set with 8' lead is available in four models. You may select a handle set for Albér Cellcorders with an intercell clamp or no intercell clamp. Figure 5. Table Of Handle Set Options summarizes these options.

Handle Set Options

Handle Set Part Number	Description
1101-005	CRT handle set without intercell clamp
1101-015	CRT handle set with intercell clamp
1101-003	CLC handle set without intercell clamp
1101-113	CLC handle set with intercell clamp

Figure 5. Table Of Handle Set Options

Push Button Testing and LEDs

LEDs on each probe light the work area and indicate test status when using the CRT. The Test push button on the red (positive) probe lets the technician start the test away from the Cellcorder.

3 Resistance Measurement Using Spike Probes

The placement of the test probes is important and is directly related to the accuracy of the readings. When using the spike probes to measure internal resistance or intercell resistance, ensure the following is done.

3.1 Make Contact To The Battery Post



CAUTION:

The contact must be to the battery post not to the mounting hardware.

When measuring internal cell resistance or intercell connector resistance, place the spike probes directly on the positive and negative terminal posts. This way, the measurement will not include any part of the external circuit. Figure 7, Figure 8, Figure 9, and Figure 10 show how to connect the spike probes for internal resistance and intercell resistance measurements.

3.2 Apply Firm Pressure

NOTE:

Firm pressure ensures that all contacts on each spike probe contact the cell being tested.

Most battery terminal posts have a layer of lead dioxide, lead sulfate, copper oxide or dirt mixed with No-Ox grease. All are high in resistance, so make sure the test probes are making good connection with the metal surfaces of the posts.

3.3 Do Not Move The Probes



WARNING:

Sparking may occur if the probes are moved during the test. Do not move the probes off the terminals while testing as there may be resultant damage to the Cellcorder.

When contact is stable, start the test. Be careful not to move or lift the probes while the test is in progress.

3.4 Inaccessible Battery Posts

On some batteries (typically VRLA modules), it is impossible to access the terminal posts because they are covered by the intercell connector. For valve regulated batteries where only the intercell and mounting bolt are accessible, place the spike probes directly on the intercell bar, as close as possible to the post. When taking readings in this manner, use Combined mode because the intercell reading will not include true strap to post connection resistance. Refer to Combined and Computed Readings on page 6.

4 CRT Resistance Testing

The red spike probe Test push button lights the LEDs on both probes and starts the resistance test. This button is similar to the Test button on the Cellcorder itself. The LEDs light the area to be tested and indicate test status. Figure 6. Table of Probe Test Functions And Conditions summarizes these features.

Probe Test Button Functions	
Press and release quickly.	Lights the LEDs on the probes for up to one minute.
Press the button for \geq half second then release.	When Cellcorder indicates Ready to Test, starts the resistance test.
On error, press for \geq half second then release.	Same as pressing the Continue function key on the Cellcorder.
On readings exist message, press for \geq a half second then release.	Same as pressing the Yes function key on the Cellcorder.
Probe LED Conditions	
LEDs are lit.	Lights up the area for connection.
LEDs blink off one time then remain lit.	Occurs when Test button is pressed. Indicates Test button may be released. Test starts.
LEDs blink off two times.	Test is successful. Also, the unit beeps when testing is complete.
LEDs blink until error indicated on Cellcorder is cleared.	Test is not successful. You may clear error using the Test button.

Figure 6. Table of Probe Test Functions And Conditions

When probe contact is stable on the posts and the Cellcorder displays Ready to Test, press the button on the red probe to start the test. After the LEDs blink off then on (about a half second), release the button. When testing is complete, the LEDs blink twice and the Cellcorder beeps. At this time, you may move the probes to test the next cell or battery. If testing is unsuccessful, usually because of intermittent post contact, the LEDs flash until the error message on the Cellcorder is cleared. Check the Cellcorder for an error message and clear when necessary.

4.1 CLC-200 Resistance Testing

The red spike probe push button starts the resistance test. This button is similar to the Test button on the CLC-200. The LEDs on the spike probes are inoperative when used with the CLC-200.

When probe contact is stable, press the button on the red probe handle for about two seconds to start the test. When testing is complete, the CLC-200 beeps. At this time, you may move the probes to test the next cell or battery. If testing is unsuccessful, usually because of intermittent post contact, clear the error message on the CLC-200 display.

5 Voltage Testing

Spike probes without an intercell clamp may be used to take float voltage readings, although you may prefer to use the smaller voltage probes supplied. Spike probes with an intercell clamp are not recommended for measuring float voltage.

When using the probes to measure float voltage, it is not necessary to press the button on the red probe handle to store the voltage reading. This will be done automatically when the Cellcorder senses a steady reading.

6 How To Make Spike Probe Connections

This chapter describes combined and computed resistance readings using probes with and without the intercell clamp. Refer to your Cellcorder User's Guide for complete descriptions of Cellcorder menus and testing.

NOTE:

Resistance testing using the Cellcorder will not begin if the unit is displaying Initializing. Wait for the indication Ready to Test before starting a resistance test.

6.1 Combined and Computed Readings

Spike probes that do not have an intercell clamp do not support intercell readings. However, they may be used in the combined or computed intercell modes to account for intercell resistance.

Combined

Spike probes can be used with the Cellcorder to produce a combined cell, plus intercell resistance reading. This one-step measurement checks the quality of the connection between the intercell bar and the posts. A Combined reading is faster than a Computed reading, but does not separately indicate cell and intercell resistance values. This type of measurement is useful when the battery posts are not accessible for taking measurements.

Computed

Spike probes can be used with the Cellcorder to produce a computed cell plus intercell resistance reading. Computed testing requires two steps. The Cellcorder displays two resistance readings: the cell R_c and the computed intercell R_{ic} . Use Computed to determine if a problem is in the cell or intercell connection. This type of measurement is useful when access to the battery does not allow placement of the intercell lead normally used for resistance testing.

6.1.1 Combined Internal Resistance

This section explains spike probe placement for combined internal resistance measurement. This is for probes without the intercell clamp.

NOTE:

Combined measurement is not available on the CLC-200.

6.1.1.1 Measuring Cell Resistance

Set the Cellcorder to Combined.

Connect the red spike probe to the positive post of the cell.

Connect the black spike probe to the negative post of the same cell.

Perform the resistance test.




Figure 7. Measuring Cell Resistance

6.1.1.2 Measuring Cell Plus Intercell Resistance

Combined testing uses two spike probes on the positive posts of two cells connected with an intercell connection.

First, set the CLC or CRT to Combined mode.

Next, connect the red spike probe to the positive post of the first cell, and connect the black spike probe to the positive post of the next cell.

Perform the resistance test.

The reading combines cell resistance R_c with the intercell resistance R_{ic} . Using the Combined mode, no micro-ohm ($\mu\Omega$) values are displayed for R_{ic1} to R_{ic4} . Combined is faster than computed, but does not separately indicate cell and intercell resistance values.

6.1.2 Computed Internal Resistance

This section shows spike probe placement for computed internal resistance measurement. This is for probes without the intercell clamp. Combined measurement is not available on the CLC-200.

6.1.2.1 Measuring Computed Cell Plus Intercell Resistance

Set the Cellcorder to Computed.

Connect the red spike probe to the positive post and connect the black spike probe to the negative post of the first cell.

Perform the resistance test.

Next, move the black spike probe to the positive post with the intercell connection of the second cell.

Perform the resistance test.




Figure 8. Combined Cell and Intercell Resistance




Figure 9. Computed Internal Resistance Step 1 & 2

The Cellcorder subtracts the lower R_c and displays two resistance readings: the cell R_c and the computed intercell R_{ic} .

6.1.2.2 Measuring Resistance with Intercell Probes

This section explains three-lead spike probe placement for simultaneous cell and intercell resistance measurement. This is for probes that have the intercell clamp. This measurement is not available on CLC-200.




Figure 10. Cell And Intercell Resistance Measurement

Determine the battery configuration and set the Cellcorder to Single, Dual, Triple or Quad accordingly. Connect the red/black intercell clamp to the negative post of the first cell. Connect the red spike probe to the positive post of the first cell. Connect the black spike probe to the positive post of the next cell. Perform the resistance test. The Cellcorder displays two resistance readings: the cell R_c and the intercell R_{ic} .

7 Index

battery	1
battery configuration	9
battery post contact.....	4
beep	5, 6
button	
test	5
caution	
contact to battery post only	4
cell	1
CLC-200.....	2
resistance testing.....	6
coaxial tip	2
computed readings.....	7
damage	
do not move probes	4
definitions.....	1
detachable probes	2
dual.....	9
external circuit.....	4
firm pressure.....	4
float voltage.....	1, 6
handle sets	3
hard to access posts	1
inaccessible	
posts	5
intercell bar.....	7
intercell resistance	4, 7
internal resistance measurement	4
jar 1	
LED.....	2, 5
measurement	
resistance	4
modular design.....	2
module.....	1
Part Number	
1100-003 CLC handle set w/o intercell clamp	3
1100-005 handle set w/o intercell clamp	3
1100-013 CLC handle set w intercell clamp	3
1100-015 handle set w intercell clamp	3

placement	
test probes.....	4
post	1, 7
posts	
inaccessible.....	5
push button testing.....	3
PVC housing.....	2
quad.....	9
Rc 9	
Rc resistance	7
readings	
computed	7
ready to test	6, 7
red spike probe	
CLC-200.....	6
resistance	
intercell measurement.....	4
internal measurement.....	4
resistance testing	
CLC-200.....	6
Ric	9
single	9
sparking	
warning	4
stable contact	5
terminal.....	1
test	
button.....	5
ready	6
test cables	1
test probes	
placement.....	4
testing	
push button	3
triple	9
VLRA	5
voltage	
float.....	1, 6
warning	
damage.....	4
do not move probes.....	4
firm pressure is necessary	4
sparking	4